



The Standards Institution of Israel

This is to certify that:

ElNet Energy & Power Quality Powermeter

Models: ElNet PQ-GR, ElNet LTC, ElNet PFC,
ElNet LT/LTE, ElNet PICO, ElNet MC,
ElNet VIP

Ratings: 100-240Vac, 50/60Hz, 20VA
Measurement circuits:
Voltage: 0-650Vac,
Max. input voltage: 1000V,
Max. input current: 6A

Manufactured by: Control Applications Ltd.


Address: 2A Habarzel St., Tel Aviv,
0069710, Israel

have been tested by the Standards Institution of Israel and
found to comply with the standard requirements of:

❖ IEC 61010-1:2010 / EN 61010-1: 2010

Safety requirements for electrical
equipment for measurement,
control, and laboratory use.
Part 1: General requirements

(Test results are detailed in SII Test Report No.: 9312314380.)



Certificate No.: 9312314380

Date of issue: 07/05/2015

Michael Terman
Head of Electrical Safety Branch
Electronics and Telematics Laboratory